



## CALIFORNIA DEVICE RECALL INFORMATION SHEET

### Philips Respironics Recalls Ventilators due to Foam Separating from the Plastic Backing That Could Sound an Alarm

Recall Date	Product Description	Recalling Firm	Recall Reason
01/18/2023	<b>Philips Trilogy 100 Ventilator</b> 20807 devices	<b>Philips Respironics, Inc.,</b> Murrysville, Pennsylvania	The replacement silicone sound abatement foam installed into some Trilogy 100 and Trilogy 200 devices may separate from the plastic backing to which it is attached. If this were to happen, the foam could potentially block air inlet, which could result in a reduction in delivered therapy volume or pressure and could cause the device to alarm.
01/18/2023	<b>Philips Trilogy 200 Ventilator</b> 537 Devices	<b>Philips Respironics, Inc.,</b> Murrysville, Pennsylvania	The replacement silicone sound abatement foam installed into some Trilogy 100 and Trilogy 200 devices may separate from the plastic backing to which it is attached. If this were to happen, the foam could potentially block air inlet, which could result in a reduction in delivered therapy volume or pressure and could cause the device to alarm.

01/18/2023	<b>Philips Garbin Ventilator</b> 1 Device	<b>Philips Respironics, Inc.,</b> Murrysville, Pennsylvania	The replacement silicone sound abatement foam installed into some Trilogy 100 and Trilogy 200 devices may separate from the plastic backing to which it is attached. If this were to happen, the foam could potentially block air inlet, which could result in a reduction in delivered therapy volume or pressure and could cause the device to alarm.
------------	--	---	---

Recall Class	Product Identification	Distribution	Affected Dates
I	<b>Philips Trilogy 100 Ventilator</b> UDI: 00606959429055 00606959429253 00606959429093 00606959429130 00606959429048 00606959429680 00606959023178 00606959429246 00606959429079 00606959429086 00606959429697 00606959429116 00606959034037 00606959429321 00606959429314 00606959429246 00606959429536 00606959429635 00606959429628 00606959429383 00606959429499 00606959429482 00606959034020 00606959429420 00606959429413	US Nationwide Distribution	December 2022 and prior
I	<b>Philips Trilogy 200 Ventilator</b> UDI: 00606959429154 00606959429208 00606959429147 00606959429185 00606959429710 00606959429611 00606959429475 00606959429260 00606959033986 00606959033993 00606959429444	US Nationwide Distribution	December 2022 and prior
I	<b>Philips Garbin Ventilator</b> UDI: 00606959429338	US Nationwide Distribution	December 2022 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)  
OR CONTACT PHILLIPS CUSTOMER SERVICE AT 800-345-6443

---

CDPH Food and Drug Branch  
MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435  
(916) 650-6500 ● (916) 650-6650 FAX  
Internet Address: [www.cdph.ca.gov](http://www.cdph.ca.gov)

