

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Premier Dental Products Co Premier Solo Diamond - Large Invented Cone

Recall Date	Product Description	Recalling Firm	Recall Reason
03/21/2025	Premier Solo Diamond - Large Invented Cone SKU: 807016C	Premier Dental Products Co	The hardness not meeting the material specification and may cause the diamond bur to bend.

Recall Class	Product Identification	Distribution	Affected Dates
II	SKU: 807016C UDI-DI: 00348783000237 Lot Numbers: M0117646, M0120260, M0123294	5 units in California	February and prior

For additional information, please visit the [FDA Website](https://www.fda.gov).