California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Pro-Med Instruments Gmbh DORO Easy-Connect Navigation Adaptor

Recall Date	Product Description	Recalling Firm	Recall Reason
05/28/2025	DORO Easy-Connect Navigation Adaptor BRAINLAB Item No.: 1204.002	Pro-Med Instruments Gmbh	Potential of compromised compatibility resulting in influence on the positioning of the reference array and an influence of the function of the navigation system cannot be excluded.

Recall Class	Product Identification	Distribution	Affected Dates
II	Model/Item No: 1204.002 UDI-DI: 04250435506738 Serial numbers: 1102010 - 2409004	1 unit in California	March and prior

For additional information, please visit the FDA Website.