

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Qiagen Sciences LLC QIAstat-Dx Respiratory Panel Plus

Recall Date	Product Description	Recalling Firm	Recall Reason
03/31/2025	Brand Name: QIAstat-Dx Product Name: QIAstat-Dx Respiratory Panel Plus Reference Number (REF): 691224 Software Version: n/a Component: n/a	Qiagen Sciences LLC	Identified faulty cartridges. If such a faulty cartridge is used for sample testing, false test results could occur.

Recall Class	Product Identification	Distribution	Affected Dates
II	REF: 691224 Lot 178038604 UDI-DI: 04053228047438 Serial numbers (SNs): 332168961 - 332169486	445 units nationwide	March and prior

For additional information, please visit the [FDA Website](https://www.fda.gov).