

## California Device Recall Information Sheet

### Food and Drug Branch – Device Recalls

Remel, Inc Zebra Thermal Barcode Printer GX430t

Recall Date	Product Description	Recalling Firm	Recall Reason
03/19/2025	Zebra Thermal Barcode Printer GX430t Model no. YSW 1000 or YSW 2000  May have been sold as an optional accessory for Sensititre Aris HIQ system YV4000-VZ	Remel, Inc	Power supply unit for Zebra Printers used with the device can potentially overheat or cause a fire hazard.

Recall Class	Product Identification	Distribution	Affected Dates
	UDI-DI: 05032384545872  <a href="#">Detailed list</a> of affected Zebra printer models  Please check the manufacturing label found either on the back or bottom of your printer. If the label lists Zebra Technologies as the manufacturer, please use the validation form on the <a href="#">Zebra website</a> to fully validate if your power supply is affected	3 units in California	February and prior

For additional information, please visit the [FDA Website](#).