

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Smith & Nephew, Inc. META-TAN LAG/Compression Screw Kit

Recall Date	Product Description	Recalling Firm	Recall Reason
05/19/2025	META-TAN LAG/Compression Screw Kit 90MM/85MM REF: 71642690	Smith & Nephew, Inc.	Integrated Lag/Compression Screw Kit has a compression screw with hexagon socket that may be undersized, which causes the screw to be non-functional, which would require replacement with another adequate size screw, and exchange may exceed 30 minutes.

Recall Class	Product Identification	Distribution	Affected Dates
II	UDI-DI: 00885556132975 Lot: 24CT81426 Expiration: 25-Mar-2034	3 units in California	March and prior

For additional information, please visit the [FDA Website](#).