California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Smiths Medical ASD, Inc. Capnocheck BCI Airway Adapter

Recall Date	Product Description	Recalling Firm	Recall Reason
03/20/2025	CAPNOCHECK BCI Airway Adapter; Model Number BSR; Product Code WW1100 CAPNOCHECK BCI Airway Adapter; Model Number BSR; Product Code WW8214	Smiths Medical ASD, Inc.	Airway adapter does not meet specification and may break off when attaching a valve or tube. Device failure may result in inaccurate readings, a break in airway circuit, and loss of ventilation. A break in the airway circuit could result in inability to ventilate, hypoxia, cardiopulmonary collapse, or death.

Recall Class	Product Identification	Distribution	Affected Dates
II	Product Code: WW1100 UDI-DI: 35019517062667 Lots: 2367264 - 4446700 Product Code: WW8214 UDI-DI 30610586038444 Lots: 022980-1-22 - 020505-1-1	9,391 units nationwide	February and prior

For additional information, please visit: FDA Website (WW1100).