

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Smiths Medical ASD, Inc. Portex Endotracheal Tube

Recall Date	Product Description	Recalling Firm	Recall Reason
03/12/2025	Portex Endotracheal Tube Item Codes 100/105/025 100/105/030 100/105/035 100/111/020 100/111/025 100/111/030 100/111/035 100/111/035 100/126/025 100/126/030 100/126/035 100/127/025 100/127/030 100/127/035 100/141/025 100/141/030 100/141/035	Smiths Medical ASD, Inc.	Affected devices have a smaller diameter than expected and may potentially result in inadequate ventilation to the patient. Patient may experience hypoxia, underdose, or cardiopulmonary collapse.

Recall Class	Product Identification	Distribution	Affected Dates
I	Item Code: 100/105/025 UDI-DI: 15019315018633 Lots: 3877711 - 4345413 Item Code: 100/105/030 UDI-DI: 15019315018640 Lots: 3880748 - 6055172 Item Code: 100/105/030 UDI-DI: 15019315018640 Lots: 3955939 Item Code: 100/105/035 UDI-DI: 15019315018657 Lots: 3874397 - 6059989	Not provided	February and prior

	Item Code: 100/105/035** UDI-DI: 15019315018657 Lots: 3955939 Item Code: 100/111/020 UDI-DI: 15019315018831 Lots: 3884360 - 6018589 Item Code: 100/111/025 UDI-DI: 15019315018848 Lots: 3874369 - 6055174 Item Code: 100/111/030 UDI-DI: 15019315018855 Lots: 3891574 - 6055175 Item Code: 100/111/035 UDI-DI: 15019315018862 Lots: 3874375 - 6055176 Item Code: 100/126/025 UDI-DI: 15019315056758 Lots: 3874389 - 6031564 Item Code: 100/126/030 UDI-DI: 15019315056765 Lots: 3891631 - 6018600 Item Code: 100/126/035 UDI-DI: 15019315056772 Lots: 3867283 - 6018590 Item Code: 100/127/025 UDI-DI: 15019315019807 Lots: 3895053 - 6055170 Item Code: 100/127/030 UDI-DI: 15019315019814 Lots: 3897978 - 6051961 Item Code: 100/127/035 UDI-DI: 15019315019821 Lots: 3874374 - 6010610 Item Code: 100/141/025 UDI-DI: 15019315019029 Lots: 3867278 - 6051980		
--	---	--	--

	Item Code: 100/141/030 UDI-DI: 15019315019036 Lots: 3872612 - 6031565 Item Code: 100/141/035 UDI-DI: 15019315019043 Lots: 3874416 - 6060009		
--	---	--	--

For additional information, please visit the [FDA Website](#).