



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Tapemark recalls Iontopatch On-the-Go Patch Therapy, Transdermal Iontophoretic Drug Delivery System, for Labeling Discrepancy**

Recall Date	Product Description	Recalling Firm	Recall Reason
11/10/2021	<p><b>iontoPATCH On-the-Go Patch Therapy</b></p> <p>Transdermal Iontophoretic Drug Delivery System, Extra Strength, 8-HOUR Average Wear Time, 6/1-patch pouches/box.            Dosage: 120mA/min</p>	<p><b>Tapemark Company</b>            Saint Paul, Minnesota</p>	<p>The dosage information on the pouch is incorrect; the primary carton label is correct.</p>

Recall Class	Product Identification	Distribution	Affected Dates
III	<p>Lot numbers:            1131-1 and 1131-2</p>	<p>3,312 cartons via Performance Health            252 cartons to other distributors, of which 28 have been returned.</p>	<p>November 2021 and prior.</p>

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch  
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435  
 (916) 650-6500 • (916) 650-6650 FAX  
 Internet Address: [www.cdph.ca.gov](http://www.cdph.ca.gov)

