

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

TruAbutment Inc. TruScan Body compatible with CONELOG 3.8

Recall Date	Product Description	Recalling Firm	Recall Reason
03/11/2025	Brand Name: TruScan Body, REF CN38-SB Product Name: TruScan Body compatible with CONELOG 3.8 Model/Catalog Number: CN38-SB Software Version: N/A	TruAbutment Inc.	Retroactive: Due to defect and quality issues dental scanner may not scan properly/accurately

Recall Class	Product Identification	Distribution	Affected Dates
II	Lot Code: KCCBB1216110 UDI: (01)08800076411791	14 units in California	December and prior

For additional information, please visit the [FDA Website](#).