

State of California—Health and Human Services Agency California Department of Public Health



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Universal Meditech Inc. Recalls Lem Fertility Hcg Pregnancy Urine Test And Lh Ovulation Test (Strips) For Inability To Continue Fulfilling Any Post-Market Responsibilities

Recall Date	Product Description	Recalling Firm	Recall Reason
11/1/2023	Lem Fertility Hcg Pregnancy Urine Test REF 100-12	UNIVERSAL MEDITECH INC. Reedley, California	Universal Meditech Inc. was violatively distributing PrestiBio" Breastmilk Alcohol Test Strip, DiagnosUSÂ; SARS-CoV-2 Antibody (IgG/IgM) Test, HealthyWiser KetoFast" Ketone Test Strips and PrestiBio" Ketone Test Strips without marketing authorization. UMI is also recall their tests because they are going out of business and would not be able to continue fulfilling any post- market responsibilities of these distributed products. UMI claims to hold 510(k)s for the other devices, purchased the intellectual property in 2015 from the previous 510(k) holder, but FDA has been unable to verify this information. The devices may have unknown performance

			characteristics because the firm is unable to provide any documentation that the devices were manufactured in conformance with the Quality System regulation, including documentation (e.g., DHF) that the distributed devices had not been modified since original clearance in a way that could impact their safety and effectiveness, documentation of controlled storage temperature/humidity , and post-market surveillance documentation, such as complaints and adverse events.
11/1/2023	Lem Fertility LH Ovulation Test (Strip) REF 200-07	UNIVERSAL MEDITECH INC. Reedley, California	Universal Meditech Inc. was violatively distributing PrestiBio" Breastmilk Alcohol Test Strip, DiagnosUS¿ SARS-CoV-2 Antibody (IgG/IgM) Test, HealthyWiser KetoFast" Ketone Test Strips and PrestiBio" Ketone Test Strips without marketing authorization. UMI is also recall their tests because they are going out of business and would not be able to continue fulfilling any post- market responsibilities of these distributed products. UMI claims to hold 510(k)s for

the other devices,
purchased the
intellectual property
in 2015 from the
previous 510(k)
holder, but FDA has
been unable to verify
this information. The
devices may have
unknown
performance
characteristics
because the firm is
unable to provide any
documentation that
the devices were
manufactured in
conformance with the
5 5
regulation, including
documentation (e.g.,
DHF) that the
distributed devices
had not been
modified since
original clearance in
a way that could
impact their safety
and effectiveness,
documentation of
Ű
temperature/humidity
, and post-market
surveillance
documentation, such
as complaints and
adverse events.
· · ·

Recall Class	Product Identification	Distribution	Affected Dates
II	Lem Fertility hCG Pregnancy Urine Test Model Number: 100-12 UDI-DI Code: None Lot Numbers: All products manufactured after March 2021 to include: hCG -S-2112-02	2500 Units Nationwide including California	22 May 2023 and prior

II Product 2 Model Number: 200-07 UDI-DI Code: None Lot Numbers: All products manufactured after March 2021 to include: LH-S-2112-01	12500 Units Nationwide including California	22 May 2023 and prior
--	---	-----------------------

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE FDA WEBSITE

CDPH Food and Drug Branch MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435 (916) 650-6500 • (916) 650-6650 FAX Internet Address: <u>www.cdph.ca.gov</u>

