California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

WHILL, INC. Recalls WHILL Powered Wheelchair

Recall Date	Product Description	Recalling Firm	Recall Reason
4/24/2026	WHILL Model C2, with software versions before HMI: 2.22 / MC: 1.13; device is an indoor/outdoor battery-operated 2-wheel drive (rear-wheel drive) powered wheelchair.	WHILL, INC. Shinagawa, Japan	Cyber vulnerability of BLE communication was found by CISA VDP program.

Recall Class	Product Identification	Distribution	Affected Dates
I	Product Code Information	4685 Units US: Kansas	December 2025 and prior

For additional information, please visit the FDA website