

California Device Recall Information Sheet

Food and Drug Branch – Device Recalls

Zimmer, Inc. NexGen LCCK Legacy Articular Surface with Locking Screw

Recall Date	Product Description	Recalling Firm	Recall Reason
01/28/2025	NexGen LCCK Legacy Articular Surface with Locking Screw - Striped Yellow/E,F - Height 10 mm Item Number: 00-5994-032-10	Zimmer, Inc.	It was identified internally that two commingle events occurred where the metal support post within the polyethylene articular surface was assembled incorrectly for three units within the scope.

Recall Class	Product Identification	Distribution	Affected Dates
II	UDI-DI: 00889024635746 Lot Numbers: 66782843 66782840 66873137	2 units in California	January and prior

For additional information, please visit the [FDA Website](#).