



California Medical Device Recall Information



Recall Name

**Abbott Recalls MitraClip® Delivery System
Due to Issues with Deploying the Clip from Delivery System**

Recall Date	Product Description	Recalling Firm	Recall Reason
02/04/16	MitraClip® Delivery System	Abbott Vascular Abbott Park, IL	<i>The company received reports involving MitraClip Delivery Systems where the user was unable to separate the implantable Clip from the delivery system.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Lot Numbers: 50811U1, 50811U2, 50812U1, 50813U1, 50814U1, 50826U1, 50826U2, 50827U1, 50908U1, 50908U2, 50909U1, 50910U1, 50910U2, 50911U1, 50924U1, 50925U1, 50928U1, 50929U1, 51010U1, 51012U1, 51012U2, 51013U1, 51014U1, 51014U2, 51026U1, 51027U3, 51028U1, 51028U2, 51029U1, 51030U1, 51105U1, 51106U1, 51109U1, 51109U2, 51110U1, 51110U2, 51117U1, 51203U1, 51204U1, 51205U1, 51207U1	CA , nationwide	Manufactured: July 14, 2015 to August 11, 2015 Distributed: August 28, 2015 to February 3, 2016

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm490774.htm>