



California Medical Device Recall Information



Recall Name

Abbott Diabetes Care Recalls Blood Glucose Monitoring Systems Due to Low Test Results

Recall Date	Product Description	Recalling Firm	Recall Reason
02/19/14	Blood Glucose Meters <ul style="list-style-type: none"> FreeStyle FreeStyle Flash 	Abbott Diabetes Care, LLC. Alameda, CA	<i>When used with the FreeStyle test strips, these meters may produce inaccurately low test results.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	All units	CA, nationwide	These two meters have not been in production since 2010.

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm389357.htm>