



# California Medical Device Recall Information



## Recall Name

**Abbott Recalls MitraClip® Delivery System  
Due to Issues with Deploying the Clip from Delivery System**

Recall Date	Product Description	Recalling Firm	Recall Reason
02/26/16	MitraClip® Delivery System	<b>Abbott Vascular</b> Abbott Park, IL	<i>The company received reports involving MitraClip Delivery Systems where the user was unable to separate the implantable Clip from the delivery system.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Lot Numbers: <b>50714U1 and greater</b>	<b>CA, nationwide</b>	Manufacturing Dates: <b>August 11, 2015 to Present</b>

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm488133.htm>