



California Medical Device Recall Information



Recall Name

**Baxter Corporation Englewood Recalls ABACUS
Total Parenteral Nutrition Calculations Software
Due to Software Errors**

Recall Date	Product Description	Recalling Firm	Recall Reason
04/02/14	ABACUS Total Parenteral Nutrition Calculations Software Versions 3.1, 3.0, 2.1, and 2.0	Baxter Corporation Englewood Englewood, CO	<i>Software failures that could result in toxic or overdose symptoms.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Product Codes: <ul style="list-style-type: none"> • 8300-0045 • 8300-0046 • 8300-0047 • 8300-0156 • 8300-0157 • 8300-0158 • 8300-0166 • 8300-0167 • 8300-0168 • 8300-0169 	Nationwide	Manufactured from May 17, 2006 through May 28, 2013 Distributed from May 17, 2006 through March 17, 2014

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm398428.htm>