



# California Medical Device Recall Information



## Recall Name

**Baxter Recalls All Unexpired Lots of 50mm 0.2 Micron Filters  
Due to the Potential for a Missing Filter Support Membrane  
And Potential Presence of Particulate Matter**

Recall Date	Product Description	Recalling Firm	Recall Reason
9/27/16	50mm 0.2 Micron Filters	<b>Baxter International, Inc.</b> Deerfield, IL	<i>The absence of a filter membrane layer and/or the presence of particulate matter in the set may contaminate a solution, which could result in bloodstream infections.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	<b>ALL Unexpired Lots</b>  Product Code: <ul style="list-style-type: none"> <li>• <b>H93835</b></li> </ul> <a href="#">Image of Product</a>	<b>CA</b> , nationwide	Distribution dates:  August 22, 2013 to June 20, 2016

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm523560.htm>