



California Medical Device Recall Information



Recall Name

Baxter International Recalls Dual Luer Lock Caps Due to Particulate Matter

Recall Date	Product Description	Recalling Firm	Recall Reason
9/19/13	<u>Dual Luer Lock Caps</u>	Baxter International, Inc. Deerfield, IL	<i>Loose particulate matter was found in product packaging. Particulate matter entering the fluid path from the Luer Lock Caps may result in thrombotic and embolic events, including pulmonary embolism, myocardial infarction and stroke.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Product Code: 2C6250 Lot # 10043 Lot # 10044	Nationwide	Distributed: June 19, 2013, and August 20, 2013

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm369253.htm>