



# California Medical Device Recall Information



## Recall Name

**BioMerieux SA Recalls NucliSENS® easyMAG® Magnetic Silica for Nucleic Acid Extraction Due to Potential for Inaccurate Test Results**

Recall Date	Product Description	Recalling Firm	Recall Reason
07/18/16	NucliSENS® easyMAG® Magnetic Silica  Reference Number: <b>280133</b>	<b>BioMerieux SA.</b> Chemin De L'Orme Marcy L'Etoile France	<i>Kits with the affected lots of the Magnetic silica have exhibited a decrease of extraction performance with certain downstream applications.</i>  <i>The detection problem could lead to a risk of false negative or invalid results for clinical laboratory tests.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Lot Numbers:  Z017BB1MS, Z017BF1MS, Z017CA1MS, Z017BE1MS, Z017CC1MS, Z017CD1MS, Z017CH1MS, Z017BD1MS, Z017DB1MS, Z017BA1MS, Z017CF1MS, Z017CE1MS, Z017CG1MS, Z017DA1MS, Z017DC1MS, Z017DD1MS, Z017EA1MS, Z017EB1MS, Z017FB1MS, Z017FA1MS, Z017KG1MS, Z017KH1MS	<b>CA</b> , nationwide	Distribution start date:  01/14/2016

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm516437.htm>