



California Medical Device Recall Information



Recall Name

Boston Scientific Recalls RotaWire Elite Guidewire and wireClip Torquer Guidewire Due to a Potential Defect

Recall Date	Product Description	Recalling Firm	Recall Reason
10/09/15	<ul style="list-style-type: none"> RotaWire Elite Guidewire wireClip Torquer Guidewire 	Boston Scientific Corp. Marlborough, MA	<i>Potential for the RotaWire 'Elite' core wires to crack and separate.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Full list of affected devices	CA, nationwide	Manufactured between: June 26, 2015 and September 10, 2015

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm474605.htm>