



California Medical Device Recall Information



Recall Name

**Philips Healthcare Recalls
Children's Medical Ventures Gel-E Donut and Squishon 2
Due to Mold Contamination**

Recall Date	Product Description	Recalling Firm	Recall Reason
11/14/14	Children's Medical Ventures: <ul style="list-style-type: none"> • Gel-E Donut gel pillow • Squishon 2 gel cushion 	Children's Medical Ventures, LLC. Monroeville, PA	<i>A potential for the presence of mold.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	All lots recalled Model Numbers: 92025-A 92025-B 92025-C 91033-2 Product Labels	CA, nationwide	Distributed between: July 2012 and August 2014

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm423199.htm>