



# California Medical Device Recall Information



## Recall Name

### Cook Medical Recalls Zilver PTX Drug Eluting Stent Due to a Possible Defect

Recall Date	Product Description	Recalling Firm	Recall Reason
4/24/13	Zilver PTX Drug Eluting Stent	<b>Cook Medical, Inc.</b> Bloomington, IN	<i>Potential for inner delivery catheter breakage.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Zilver PTX Drug Eluting Stent  Suspected Catalog Number Recalled: <ul style="list-style-type: none"> <li>ZIV6*****PTX</li> </ul> Recall includes all sizes, diameters and lot numbers  <a href="#">Product Labels</a>	CA, global	Distributed between December 13, 2012 and April 16, 2013

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm349421.htm?source=govdelivery>