



California Medical Device Recall Information



Recall Name

Draeger Recalls Fabius Anesthesia Machines Due To a Potential Defect

Recall Date	Product Description	Recalling Firm	Recall Reason
8/14/13	Fabius Anesthesia Machines	Draeger Medical, Inc. Telford, PA	<i>Potential for a failure of the automatic ventilation function of the device.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Models of Anesthesia Machines recalled: <ul style="list-style-type: none"> • Fabius GS Premium • Fabius OS • Fabius Tiro • Fabius Tiro D-M List of Affected Models and Serial numbers	CA , nationwide	Distributed between: March 2013 and June 2013

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm370834.htm>