



California Medical Device Recall Information



Recall Name

**DePuy Synthes Recalls Power Tool System Battery Adaptors
Due to Possible Explosion Risk**

Recall Date	Product Description	Recalling Firm	Recall Reason
6/24/16	Adaptor and Light Adaptor for <ul style="list-style-type: none"> • Small Battery Drive • Small Battery Drive II 	DePuy Synthes, Palm Beach Gardens, FL	<i>Potential for the adaptors to produce extreme internal pressure, which may cause the device to explode.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Suspect S/Ns recalled: <ul style="list-style-type: none"> • 05.001.024 • 05.001.108 	CA , nationwide	Distribution Dates: January 2006 to June 2016

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm522949.htm>