



California Medical Device Recall Information



Recall Name

Diabetic Supply of Suncoast Recalls

Advocate Redi-Code+ Blood Glucose Test Strips Due to a Labeling Error

Recall Date	Product Description	Recalling Firm	Recall Reason
6/06/14	Advocate Redi-Code+ Blood Glucose Test Strips BMB-BA006A	Diabetic Supply of Suncoast, Inc. Dorado, PR	<i>Labeling error (omission of test strip model number) could result in confusion about which meter models the Redi-Code+ blood glucose test strips are designed to be used with.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	All lots recalled Product Labels	CA , nationwide	Distributed between: January 2013 and June 6, 2014

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm400559.htm>