



California Medical Device Recall Information



Recall Name

Fisher and Paykel Recalls Infant Nasal CPAP Prongs Due to Potential Detachment from Nasal Tubing

Recall Date	Product Description	Recalling Firm	Recall Reason
04/15/14	<ul style="list-style-type: none"> • Infant Nasal CPAP Prongs (10 Pack), and • Bubble CPAP Starter Kits 	Fisher and Paykel Healthcare, Ltd. Auckland, New Zealand	<p><i>The firm received reports of the affected prongs detaching from the nasal tubing during use.</i></p> <p><i>When the affected prongs detach from the nasal tubing, therapy is likely to be interrupted.</i></p>
Recall Class	Product Identification	Distribution	Affected Dates
I	Affected model numbers and lot numbers.	CA, nationwide	<p>Manufactured and distributed from</p> <p>June 6, 2013 through March 25, 2014.</p>

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm398631.htm>