



California Medical Device Recall Information



Recall Name

GE Healthcare Recalls Magnetic Resonance Imaging (MRI) System with Magnet Rundown Unit Due to a Possible Delay in Shutting Off the Magnetic Field

Recall Date	Product Description	Recalling Firm	Recall Reason
1/12/15	Magnetic Resonance Imaging (MRI) Systems - with Magnet Rundown Unit (MRU)	GE Healthcare Waukesha, WI	<i>Potential for a disconnected MRU.</i> <i>A disconnected MRU could delay shut off of the MRI in case of an emergency, such as when a metal (ferrous) object is brought into the magnetic field.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	ALL GE Healthcare MRI systems with superconducting magnets Link to affected MRI systems Serial Numbers	CA , nationwide	Manufactured and distributed from: 1985 through December 2014

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm435537.htm>