



California Medical Device Recall Information



Recall Name

**Hospira Recalls GemStar Power Supply, 3VDC
Due to Potential for Malfunction**

Recall Date	Product Description	Recalling Firm	Recall Reason
02/03/14	GemStar Power Supply, 3VDC <ul style="list-style-type: none"> • Wall Mount Power Supply List Number 13026-05 • Desk Top Power Supply List Number 13072-05 	Hospira, Inc. Lake Forest, IL	<i>The power supply may not properly deliver electric power to the GemStar Infusion Pump, which has been found operating on battery power while connected to the 3VDC power supply.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Affected Lot Numbers	CA, nationwide	Distributed (U.S.A.): <u>List 13072-04-05</u> Nov. 2011 to April 2013 <u>List 13026-04-05</u> Feb. 2013 to April 2013

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm421946.htm>