



California Medical Device Recall Information



Recall Name

ICU Medical Recalls ConMed Stat2 Flow Controller Due to Potential of Higher Flow Rates

Recall Date	Product Description	Recalling Firm	Recall Reason
8/19/14	ConMed Stat2 Flow Controller Item Numbers: <ul style="list-style-type: none"> • 011-C9801 • 011-C9801 • AH7007 • B9897 • Z2648 	ICU Medical, Inc. , San Clemente, CA	<i>Potential for the flow controller to deliver fluid at a much higher flow rate than what is set.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Affected Lot Numbers: 2768416 2768417 2758229 2785379 2801951 34-128-HE 34-540-Y1 35-151-SJ 35-805-JW 36-137-SL 36-469-SL	CA , nationwide	Distributed from: October 2013 to January 2014

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm418109.htm>