



California Medical Device Recall Information



Recall Name

I-Flow Corporation Recalls ON-Q Pump with ONDEMAND Bolus Button Due to Possible Malfunction

Recall Date	Product Description	Recalling Firm	Recall Reason
5/08/12	ON-Q Pump with ONDEMAND Bolus Button	I-Flow Corporation Lake Forest, CA	<i>Bolus Button may not lock in the down position</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	ON-Q Pump with ONDEMAND Bolus Button Implicated Lots Recalled: <ul style="list-style-type: none"> All lots with a 10-digit lot number sequentially less than and including lot # "0200521454" 	CA , nationwide	Manufactured from October 18, 2002 through April 30, 2012

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm317826.htm>