



California Medical Device Recall Information



Recall Name

LifeScan Recalls OneTouch Verio IQ Blood Glucose Meters Due to Units Shutting Off at High Blood Glucose Levels

Recall Date	Product Description	Recalling Firm	Recall Reason
03/11/13	All OneTouch Verio IQ Blood Glucose Meters	LifeScan, Inc. Milpitas, CA	<i>At extremely high blood glucose levels (1024 mg/dL and above), the meter will not provide a warning and will shut off.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	All OneTouch Verio IQ Blood Glucose Meters	CA , nationwide	Distributed 12/14/11 to 03/07/13.

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm349127.htm>