



# California Medical Device Recall Information



## Recall Name

### Maquet Cardiovascular Recalls SERVO-i Battery Modules Due to Short Battery Life

Recall Date	Product Description	Recalling Firm	Recall Reason
3/07/13	Battery Modules: <ul style="list-style-type: none"> <li>NiMH battery;</li> <li>For use with <b>Maquet SERVO-i Ventilator Systems</b>.</li> </ul>	<b>Maquet Cardiovascular</b> Wayne, NJ	<i>Battery modules have shorter run time than expected which could result in unexpected ventilator shut downs leading to potential adverse health consequences and/or death.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	SERVO-i battery modules: <ul style="list-style-type: none"> <li>Part No. <b>6487180</b></li> <li>Date codes (identified on the battery's label):  <b>"1005" thru "1243"</b></li> </ul>	<b>CA</b> , nationwide	Distributed between Feb. 15, 2010 and Nov. 19, 2012.

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm351092.htm>

<http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRes/res.cfm?ID=116703>