



California Medical Device Recall Information



Recall Name

**Nurse Assist Recalls All Unexpired Lots of I.V. Flush Syringes
Due to a Potential Bacterial Contamination**

Recall Date	Product Description	Recalling Firm	Recall Reason
10/04/16	I.V. Flush Syringes	Nurse Assist, Inc. , Haltom City, TX	<i>Potential link to <u>Burkholderia cepacia</u> bloodstream infections with the product.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Suspect Product Codes: <ul style="list-style-type: none"> • 1203 (3ml fill) • 1205 (5ml fill) • 1210 (10ml fill) • 1210-BP (10ml fill) ALL Unexpired Lots	CA , nationwide	Distributed between: February 16, 2016 and September 30, 2016

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm523959.htm>