



California Medical Device Recall Information



Recall Name

Nephros Recalls SafeSpout / SafeShower Water Filtration Products Due to Possible Exposure to Harmful Bacteria

| Recall Date | Product Description | Recalling Firm | Recall Reason |
|--------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------|--------------------------------------------------------------------------------------------------------------------------|
| 10/28/13 | <ul style="list-style-type: none"> SafeSpout filters SafeShower Hand Held (HH) filters SafeShower Fixed Head (FH) filters | Nephros, Inc. River Edge, NJ | <i>Potential for the fiber filter or the sealing compound to break apart and expose patients to bacteria or viruses.</i> |
| Recall Class | Product Identification | Distribution | Affected Dates |
| I | SafeSpout Item Numbers: <ul style="list-style-type: none"> 70-0233 70-0238 SafeShower HH Item Number: <ul style="list-style-type: none"> 70-0237 SafeShower FH Item Number: <ul style="list-style-type: none"> 70-0236 All Lots Recalled | CA , nationwide | Distributed from: October 2011 to September 2013 |

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/medicaldevices/safety/listofrecalls/ucm380452.htm>