



California Medical Device Recall Information



Recall Name

Novo Nordisk Recalls GlucaGen HypoKit (glucagon [rDNA origin] for injection)

Due to Detached Needles on the Syringe in the Kit

Recall Date	Product Description	Recalling Firm	Recall Reason
09/08/16	GlucaGen HypoKit (glucagon [rDNA origin] for injection)	Novo Nordisk, Inc., Plainsboro, NJ	<i>Potential that needle could be detached from the syringe containing Sterile Water for Injection.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Suspect recalled batches: <ul style="list-style-type: none"> • FS6X270 • FS6X296 • FS6X538 • FS6X597 • FS6X797 • FS6X875 Expiry: 09/30/2017	CA, nationwide	Distributed starting: February 15, 2015

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm519872.htm>