California Medical Device Recall Information

Recall Name
Siemens Recalls MicroScan Synergies plus and MicroScan rapID/S plus Negative Panels Due to False Susceptible and False Intermediate Results

Recall Date	Product Description	Recalling Firm	Recall Reason
8/21/13	MicroScan Synergies plus® MicroScan® rapID/S plus Negative Panels	**Siemens Healthcare Diagnostics, Inc.** West Sacramento, CA	*Potential to report false susceptible and false intermediate results of imipenem and meropenem antimicrobial susceptibility testing.*
Recall Class	**Product Identification**	**Distribution**	**Affected Dates**
I	Siemens Material Number / Part Number: • Synergies plus Negative Urine Combo 1 - 10444745 / B1025-106 • Synergies plus Negative Combo 2 - 10444747 / B1025-108 • Synergies plus Negative Breakpoint Combo 7 - 10444748 / B1025-109 • Synergies plus Negative Urine Combo 2 - 10444749 / B1025-112 • Synergies plus Negative Urine Combo 5 - 10483101 / B1025-115	**CA**, nationwide	Distributed between: 07/11/11 and 08/02/13

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm368112.htm