



California Medical Device Recall Information



Recall Name

Symbios Medical Products Recalls GoPump and GoBlock Kits Due to Excessively High Flow Rates

Recall Date	Product Description	Recalling Firm	Recall Reason
05/10/13	<ul style="list-style-type: none"> GoPump Rapid Recovery System Kits GoBlock Kits <p>Product Labels</p>	Symbios Medical Products, LLC Indianapolis, IN	<i>Potential for excessively high flow rates resulting in rapid influx of medication, which presents a risk of patient toxicity and serious injury.</i>
Recall Class	Product Identification	Distribution	Affected Dates
TBD	See the Product List below for GoPump and GoBlock Kits subject to the recall. <p>Product List</p>	CA , nationwide	Distributed between 04/01/11 and 04/30/13.

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm352627.htm?source=govdelivery>