



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Abbott Medical Recalls Infinity 7 Implantable Pulse Generator Ref 6662 For Complaints From Patients Who Are Unable To Exit Mri (Magnetic Resonance Imaging) Mode On Their Ipgs

Recall Date	Product Description	Recalling Firm	Recall Reason
9/20/2023	Abbott Medical Infinity 7 Implantable Pulse Generator Ref 6662	Abbott Medical Plano, Texas	Firm has received complaints from patients who are unable to exit MRI (Magnetic Resonance Imaging) mode on their IPGs

Recall Class	Product Identification	Distribution	Affected Dates
I	Abbott Medical Infinity 7 Implantable Pulse Generator Ref 6662 Model Number: 6662 UDI-DI Codes: 05415067030023	7,793 Units Nationwide including California	July 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

