



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Alphatec Spine Recalls Lif Amp For Adjustable Awl

Recall Date	Product Description	Recalling Firm	Recall Reason
9/6/2023	Lif Amp Adjustable Awl, REF 117-165, Part of the AMP System. Used with Spine Lateral Interbody Systems, IdentiTi LIF, Transcend LIF and Battalion LLIF,	ALPHATEC SPINE, INC. Carlsbad, California	Awl instrument adjustable drill button assembly assembled in the incorrect orientation, preventing locking feature from engaging with shaft at the desired set point; instruments lack ability to control awl depth by means of adjustable shaft advancement stop, which may lead to over insertion, dural tear, vascular/neurologic injury, adjacent tissue damage, increased operative time, revision surgery.

Recall Class	Product Identification	Distribution	Affected Dates
II	LIF AMP UDI-DI: 00190376228037, Lot: EM50715	29 Units Nationwide including California	July 7, 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

