



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Avante Health Solutions Philips Respironics V60 Ventilator for Device Contamination

Recall Date	Product Description	Recalling Firm	Recall Reason
7/5/2023	Philips Respironics V60 Ventilator Model Number: R1053618 and R1076709	Avante Health Solutions Louisville, KY	Devices were imported with Japanese Language and NIST Oxygen Gas Fittings and retrofitted, may pose a serious health risk as the gas pathway of these devices may be contaminated.

Recall Class	Product Identification	Distribution	Affected Dates
II	Philips Respironics V60 Ventilator USI-DI: 00884838009868 Serial Numbers: Serial Number 100043954 100050078 100052760 100052848 100054619 100054626 100055277 100055557 100058549 100058566 100058837 100058995 100060228 100065373 100069520 100070487 100071122 100072691 100078379 100080232 100080844 201003462	2 Units in California	May 2023 and Prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

