



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**B. Braun Medical Recalls Perifix<sub>2</sub>/Epidural Anesthesia Set for Incorrect Lid Stock Label**

Recall Date	Product Description	Recalling Firm	Recall Reason
9/27/2023	<b>Perifix<sub>2</sub> / Epidural Anesthesia Set</b> Medicated	<b>B. BRAUN MEDICAL, INC.</b> Allentown, Pennsylvania	Product may be mislabeled with an incorrect lid stock label.

Recall Class	Product Identification	Distribution	Affected Dates
II	<b>Perifix<sub>2</sub> / Epidural Anesthesia Set</b> UDI-DI: 04046964178726 Product Code: 332250; Lot Number 0061863524, expiration date 30.APR.2025	3440 Units Nationwide including California	31 July 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

