



CALIFORNIA DEVICE RECALL INFORMATION SHEET

Baxter Healthcare Corporation for an Increase in Reported False Upstream Occlusion Alarms

Recall Date	Product Description	Recalling Firm	Recall Reason
7/19/2023	SIGMA Spectrum Infusion Pump Product Code 35700BAX2	Baxter Healthcare Corporation Deerfield, IL	A medical device correction is being issued for SIGMA Spectrum Infusion System (V8 Platform) and Spectrum IQ Infusion System with Dose IQ Safety Software infusion pumps due to an increase in reported false upstream occlusion alarms following upgrades to software versions v8.01.01 and v9.02.01.
7/19/2023	The Spectrum IQ Infusion System with Dose IQ Safety Software Product Code 3570009	Baxter Healthcare Corporation Deerfield, Illinois	An increase in reported false upstream occlusion alarms following upgrades to software versions v8.01.01 and v9.02.01

Recall Class	Product Identification	Distribution	Affected Dates
I	SIGMA Spectrum Infusion Pump GTIN 00085412498683, Software version v8.01.01, Serial Numbers	3306 Units Nationwide	April 2023 and Prior

I	The Spectrum IQ Infusion System with Dose IQ Safety Software UDI/DI 00085412610900, Software Version v9.02.01, All Serial Numbers	19861 Units Nationwide	April 2023 and Prior
---	---	---------------------------	----------------------

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch
MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435
(916) 650-6500 • (916) 650-6650 FAX
Internet Address: www.cdph.ca.gov

