



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Bearcare Recalls Walnut Wearble Smart Thermometer Because Of Skin Burns To Children**

Recall Date	Product Description	Recalling Firm	Recall Reason
6/15/2023	<b>Walnut Wearable Smart Thermometer</b> Intended for continuous chest temperature monitoring of children ages 0-6 years in non-emergency medical situations, Model Number WT20	<b>BEARCARE, INC.</b> Hauppauge, New York	There have been reports of skin burns to children wearing the device.

Recall Class	Product Identification	Distribution	Affected Dates
I	<b>Walnut Wearable Smart Thermometer</b> Lot Number 20221115W002	1818 Units Nationwide	May 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch  
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435  
 (916) 650-6500 • (916) 650-6650 FAX  
 Internet Address: [www.cdph.ca.gov](http://www.cdph.ca.gov)

