



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Vyaire Medical Recalls Bellavista 100 Ventilator for Potential Cessation of Ventilation in Certain Circumstances**

Recall Date	Product Description	Recalling Firm	Recall Reason
12/20/2021	<p><b>bellavista 1000/1000e ventilator</b>            Catalog #301.100.030 and 1000e , 301.100.130.</p> <p>The bellavista 1000/1000e ventilator is intended to provide positive pressure ventilatory support to adult and pediatric patients, and optionally, infant and neonatal patients.</p>	<b>Vyaire Medical</b> Mettawa, Illinois	Potential cessation in ventilation can occur under specific conditions

Recall Class	Product Identification	Distribution	Affected Dates
N/A	All Lot numbers	Nationwide including California	December 2021 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch  
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435  
 (916) 650-6500 • (916) 650-6650 FAX  
 Internet Address: [www.cdph.ca.gov](http://www.cdph.ca.gov)

