



CALIFORNIA DEVICE RECALL INFORMATION SHEET

Boston Scientific Neuromodulation Corporation Superior Indirect Decompression System

Recall Date	Product Description	Recalling Firm	Recall Reason
6/21/2023	Superion Indirect Decompression System Driver Instrument, part of the Superior Indirect Decompression System, REF: 102-9800, used with the Superior Indirect Decompression System IFU, Superior IDS Kit IFU, Surgical Technique Manual	Boston Scientific Neuromodulation Corporation Valencia, CA	Pending update to indirect decompression system instructions for use informing users that excessive force during the implant procedure may cause driver instrument tip breaks, which may result in metal fragments (Driver teeth/tips) within the implant location; and if metal fragments are not removed and remain in situ, MRI scans are NOT advised due to potential risk of patient injury.

Recall Class	Product Identification	Distribution	Affected Dates
II	Superion Indirect Decompression System Driver UDI-DI: 00884662000574, UPN: 102-9800, All Lots. Superior Indirect Decompression System IFU (92479815-02), Superior IDS Kit IFU (92479820-02), Surgical Technique Manual (92479821-02)	2,570 Units in California	May 2023 and Prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch
MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435
(916) 650-6500 ● (916) 650-6650 FAX
Internet Address: www.cdph.ca.gov

