



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Draeger Medical Recalls Anti-Asphyxia Valve and Standard Elbow for NovaStar TS due to Potential Impact of Magnets on Medical Devices and Metallic Implants

Recall Date	Product Description	Recalling Firm	Recall Reason
3/1/2023	NovaStar TS, Anti-Asphyxia Valve Reusable, Material Numbers MP01576 (Size S), MP01577 (Size M), and MP01578 (Size L)	Draeger Medical Inc. Telford, Pennsylvania	The magnets used in the swivel frame/tube set and the head-gear clips might impact devices (implanted and not) and metallic implants.
3/1/2023	NovaStar TS, Standard elbow Reusable, Material Numbers MP01579 (Size S), MP01580 (Size M), and MP01581 (Size L)	Draeger Medical Inc. Telford, Pennsylvania	The magnets used might impact medical devices & implants.

Recall Class	Product Identification	Distribution	Affected Dates
I	Anti-Asphyxia Valve All Lot numbers. UDI-DI: 04048675041146 (MP10576); 04048675041153 (MP01577); 04048675041160 (MP01578)	16 units nationwide	January 2023 and prior
I	Standard Elbow All Lot numbers. UDI-DI: 04048675040552 (MP01579); 04048675040569 (MP01580); 04048675040538 (MP01581)	893 units nationwide	January 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

