



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Eitan Medical Ltd Recalls Sapphire Infusion Pumps For Failure To Detect Air In The Line

Recall Date	Product Description	Recalling Firm	Recall Reason
12/1/2023	Sapphire Multi-Therapy Infusion Pump REF 15031-000-0028	EITAN MEDICAL LTD Netanya, Israel	Software issues in software version Rev 16.10. Pumps may fail to detect air in the line when runn-ing on battery power.
12/1/2023	Sapphire Epidural Infusion Pump REF 15032-000-0027	EITAN MEDICAL LTD Netanya, Israel	Failure to detect air in the line.
12/1/2023	Sapphire Plus Infusion Pump REF 15038-000-0001	EITAN MEDICAL LTD Netanya, Israel	Failure to detect air in the line.

Recall Class	Product Identification	Distribution	Affected Dates
I	Sapphire Multi-Therapy Infusion Pump UDI-DI: 15031-000-0028/7290109150109	1383 Total Units Nationwide	September 2023 and prior
I	Sapphire Epidural Infusion Pump UDI-DI: 15032-000-0027/7290109150147	1383 Total Units Nationwide	September 2023 and prior
I	Sapphire Plus Infusion Pump UDI-DI: 15038-000-0001/7290109150161	1383 Total Units Nationwide	September 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#) or email [Eitan Medical Ltd](#)

