



State of California—Health and Human Services Agency  
California Department of Public Health



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Full Vision Inc Trackmaster Treadmill for Treadmill Malfunction**

Recall Date	Product Description	Recalling Firm	Recall Reason
7/12/2023	<b>TRACKMASTER TMX428 110V</b> MODEL NUMBER 317-07926	<b>Full Vision Inc</b> Newton, KS	Treadmill malfunction: drive PCB can misread the signals and stop or reverse and then accelerate forward.
7/12/2023	<b>TRACKMASTER TMX428 220V</b> MODEL NUMBER 317-07927	<b>Full Vision Inc</b> Newton, KS	Treadmill malfunction
7/12/2023	<b>TRACKMASTER TMX428CP</b> MODEL NUMBER 317-07928	<b>Full Vision Inc</b> Newton, KS	Treadmill malfunction
7/12/2023	<b>TRACKMASTER TMX428CP 220V</b> MODEL NUMBER 317-07929	<b>Full Vision Inc</b> Newton, KS	Treadmill malfunction
7/12/2023	<b>TRACKMASTER GE T2100-ST1 110VV</b> MODEL NUMBER 317-07926GE	<b>Full Vision Inc</b> Newton, KS	Treadmill malfunction
7/12/2023	<b>TRACKMASTER GE T2100-ST2 220V</b> MODEL NUMBER 317-07927GE	<b>Full Vision Inc</b> Newton, KS	Treadmill malfunction

<b>Recall Class</b>	<b>Product Identification</b>	<b>Distribution</b>	<b>Affected Dates</b>
II	<b>TRACKMASTER TMX428 110V</b> UDI/DI 00860176000606, Serial Numbers: FVDC-7585- FVDC-8869...	20 Units in California	May 2023 and Prior
II	<b>TRACKMASTER TMX428 220V</b> UDI/DI 00860176000613, Serial Numbers: FVDC-7630- FVDC-8808...	11 Units in California	May 2023 and Prior
II	<b>TRACKMASTER TMX428CP</b> UDI/DI 00860176000620, Serial Numbers: FVDC-7644- FVDC-8816...	5 Units in California	May 2023 and Prior
II	<b>TRACKMASTER TMX428CP 220V</b> UDI/DI 00860176000637, Serial Numbers: FVDC-7641- FVDC-8888...	51 Units in California	May 2023 and Prior
II	<b>TRACKMASTER GE T2100-ST1 110VV</b> UDI/DI 00860176000668, Serial Numbers: GEDC-6632- GEDC-8347...	52 Units in California	May 2023 and Prior
II	<b>TRACKMASTER GE T2100-ST2 220V</b> UDI/DI 00860176000675, Serial Numbers: GEDC-6608- GEDC-8382...	20 Units in California	May 2023 and Prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)