



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Haimen Shengbang Laboratory Equipment Recalls Viral Transport Container**

Recall Date	Product Description	Recalling Firm	Recall Reason
7/12/2022	<b>HXBL/WETEX/Genesis Biomedical Viral Transport Container</b> Model No. HX-K59 Activated Type Model No. HX-K59-1 Custom Made	<b>Haimen Shengbang Laboratory Equipment Co. Ltd.</b> Nantong, China	Distributed VTM outside of VTM Guidance and without clearance.

Recall Class	Product Identification	Distribution	Affected Dates
I	<b>HXBL/WETEX/Genesis Biomedical Viral Transport Container</b> Model No. HX-K59 Activated Type No UDI All Lots	8,710,600 total units Nationwide, including California	July 2022 and prior
I	<b>HXBL/WETEX/Genesis Biomedical Viral Transport Container</b> Model No. HX-K59-1 No UDI All Lots	3,950,000 total units Nationwide, including California	July 2022 and prior
I	<b>HXBL/WETEX/Genesis Biomedical Viral Transport Container</b> Custom Made No UDI All Lots	100,000 total units Nationwide, including California	July 2022 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

